



# INDIAN TESTING CENTRE (P) LTD.

101, SASHCO BHAWAN, NEAR AAKASH CINEMA, AZADPUR COMMERCIAL COMPLEX  
AZADPUR, DELHI-110033 (INDIA). TELEFAX : 91-11-27671535, 27671542, 24505822  
visit us at www.indiantestingcentre.org E-mail : info@indiantestingcentre.org



T-1699/T-1699  
CQR/004(00/01)

ISSUED TO:- The Principal,  
DAV Centenary Public School,  
Sector B-4, Pocket - 5, DDA Project,  
Narela, Delhi - 110040.

TEST CERTIFICATE NO.:- 6202506126001  
DATE:- 19 - JUNE - 2025  
REFERENCE NO. - DAVCPS/405/2025-26  
DATE OF RECEIPT:- 12 - JUNE - 2025  
PAGE:- 1 OF 1

SAMPLE DESCRIPTION :- One Sample of Drinking Water was Submitted.

## TEST CERTIFICATE

S.NO.	TESTS	TEST METHODS	RESULTS	REQUIREMENTS ( IS : 10500 - 2012 )	
				ACCEPTABLE LIMIT	PERMISSIBLE LIMIT IN ABSENCE OF ALTERNATE SOURCE
1.	Colour, Hazen Unit	IS:3025 (Pt.-4)	< 1	5.00 Max.	15.00 Max.
2.	Odour	IS:3025(Pt.-5)	agreeable	Agreeable	Agreeable
3.	Turbidity, NTU	IS:3025(Pt.-10)	< 1	1.00 Max.	5.00 Max.
4.	Ph Value	IS:3025(Pt.-11)	7.25	6.50 - 8.50	No Relaxation
5.	Total Hardness as CaCO <sub>3</sub> , mg/l	IS:3025(Pt.-21)	81.00	200.00 Max.	600.00 Max.
6.	Iron as Fe, mg/l	IS:3025(Pt.-53)	0.10	0.30 Max.	No Relaxation
7.	Chloride as Cl, mg/l	IS:3025(Pt.-32)	48.50	250.00 Max.	1000.00 Max.
8.	Total Dissolved Solids, mg/l	IS:3025(Pt.-16)	151.00	500.00 Max.	2000.00 Max.
9.	Calcium as Ca, mg/l	IS:3025(Pt.-40)	9.50	75.00 Max.	200.00 Max.
10.	Sulphate as SO <sub>4</sub> , mg/l	IS:3025(Pt.-24)	32.00	200.00 Max.	400.00 Max.
11.	Total Alkalinity as CaCO <sub>3</sub> , mg/l	IS:3025(Pt.-23)	27.00	200.00 Max.	600.00 Max.
12.	Magnesium as Mg, mg/l	IS:3025(Pt.-46)	5.90	30.00 Max.	100.00 Max.
13.	Fluoride as F, mg/l	IS:3025(Pt.-60)	< 1	1.00 Max.	1.50 Max.
14.	Nitrate as NO <sub>3</sub> , mg/l	IS:3025(Pt.-34)	1.00	45.00 Max.	No Relaxation
15.	E. Coli per 100 ml	IS:15185-2016	absent	absent/100 ml	--
16.	Coliform per 100 ml	IS:15185-2016	absent	absent/100 ml	--

REMARKS:- The sample conforms to the acceptable limits of IS : 10500 - 2012 for drinking purpose with respect to above tests.

\*\*\*\*\*



AUTHORISED SIGNATORY